



शासकीय वैद्यकीय महाविद्यालय, चंद्रपूर

Government Medical College, Chandrapur

Govt. T.B. Hospital Campus, Ramnagar, Chandrapur 442401 M.S.

Telephone No. 07172- 277103, 277106, 252522 Website: www.gmccchandrapur.org Email: gmccchandrapur@gmail.com

NO.GMCC/CENTRAL STORE/ Lab.Reagent Refrigerator/ 4388 /24

Dt./0 /06/ 2024

Open Quotation Notice (Website) For the supply of Laboratory Reagent Refrigerator

Dean, Govt. Medical College, Chandrapur invites open quotation for the supply of Laboratory Reagent Refrigerator Government Medical College, Chandrapur from Manufacturer/Dealers. Detail list & specification is enclosed. Interested supplier/Manufacturer/dealer can send their quotation with their authorization certificate attach with this quotation.

Quotation should be sent in sealed envelope (duly sealed by selling wax) by post or hand delivery so as to reach this office on or before 21 / 06 /2024 up to 5.30 pm
Date & timing of the opening of quotation, - 24 / 06 /2024 at 04:00 pm. at College Council Hall.
Representative of the supplier may attend the meeting by submitting authorization letter.


1. The word "Quotation for Laboratory Reagent Refrigerator" last Date. 21 /06/2024 should be clearly written on cover.
2. Quotation, which received late, will not be accepted under any circumstances.
3. Rate should be for free delivery at Govt. Medical College, Chandrapur premises only. Rate should be quoted including all charges. Rates must be quote on letter head along with Pan & GST Number.
(Certificate must be attached herewith and Gumastha Certificate including Shop Registration must be attached)
4. Quoted goods should be strictly accordingly to the specification mentioned in the list. Make Model of items should be specifically stated in quotation and catalogue/leaflets etc. should be submit along with the quotation.
5. Catalogue/Leaflet is must along with quotation received and the price of the same will be considered on leaflets/catalogue.
6. The concern head of the department can ask for demo of the lowest price model and it will be binding for the supplier to give the demo in stipulated period (i.e. within 4 days) at Chandrapur. If the supplier fails to give the demonstration of the same model in stipulated period then his quotation will be treated as cancelled.
7. After placing the order to lowest one, the order will have to execute in full within the stipulated time. If the supplier fails to comply, within stipulated period order may be cancelled and he will be declared defaulter and his any offer will not be consider in future.

8. The undersigned reserved the rights to accept or reject any or all quotation without giving any reason.
9. The quotation should be sent in the name of the **Dean, Govt. Medical College, Chandrapur.**
10. Payment of the order goods will be made by CMP. For the CMP purpose (unregistered) supplier should submit copy of PAN card, cancelled cheque, one photograph & bank details.
11. The supplier should deliver the Equipment in the Surgery department of the institution.


Laboratory Reagent Refrigerator

Sr. No.	Description of Item	Per Qty.	Price Per Equipment (Including GST)
01	Reagent Refrigerator	1 Nos	
Technical Specification as per Annexure 'A'			

Attached :- Annexure 'A'


Dean

Govt. Medical College
Chandrapur



New range of refrigerators from REMI are designed for storage applications in pharmaceutical, microbiology, biotechnology, chemical, clinical diagnostic, hospitals & research institutions. The unit is supplied with a glass door with key lock for convenient sample viewing, forced air circulation system ensures uniform temperature throughout the chambers. Microprocessor controller is provided for display of internal temperature along with alarm systems.

Salient Features :

- Inner & Outer body pre painted G. I.
- Energy efficient CFC free refrigeration
- Uniform cooling by forced air circulation & frost free
- Adjustable Plastic coated wire mesh shelves
- Clear product visibility with triple glass
- High / low temperature alarm
- Automatic Defrosting
- Alarms: High/Low temperature variation
- Auto door closure
- Built in voltage stabilizer

Optional :

- Battery backup for temperature display & alarm for over 12 Hrs

Technical Data				
Model No.	RLR-200	RLR-300	RLR-400	RLR-500
Capacity (Litres)	200	300	400	500
Temperature Range	1° C ~ 10° C (Uniformity 2° C) ±			
No. of Door (Glass)	1	1	1	1
Baskets / Shelves	3	4	4	4
External Dimensions (mm)				
Width	500	585	583	640
Depth	660	715	715	785
Height	1660	1792	2053	2135

Supply: 220-240 Volts 50 Hz Single Phase.

[Signature]
 06/06/2024
 Head
 Department of Biochemistry
 Govt. Medical College,